

# *Allosteric effects in haemoglobin*

KIYOHIRO IMAI

*Department of Physicochemical Physiology, Medical School  
Osaka University, Nakanoshima, Osaka*

C A M B R I D G E U N I V E R S I T Y P R E S S

*Cambridge*

*London New York New Rochelle*

*Melbourne Sydney*

**EXHIBIT**

D

*GP*  
*96.2*  
*T47*

Published by the Press Syndicate of the University of Cambridge  
The Pitt Building, Trumpington Street, Cambridge CB2 1RP  
32 East 57th Street, New York, NY 10022, USA  
296 Beaconsfield Parade, Middle Park, Melbourne 3206, Australia

© Cambridge University Press 1982

First published 1982

Printed in Great Britain at the University Press, Cambridge

Library of Congress catalogue card number: 81-21581

*British Library Cataloguing in Publication Data*

Imai, Kiyohiro  
Allosteric effects in haemoglobin.

1. Haemoglobinopathy
  1. Title
- 616.1'51 RC641.7.H35  
ISBN 0 521 22575 2

absorption spectrum in the visible region or in the near ultraviolet region (Soret region) is recorded (Fig. 4.1). A small volume of air is injected into the tonometer, the liquid and gas phases are equilibrated at a constant temperature, and another spectrum is recorded. This procedure is repeated until the haemoglobin is fully oxygenated.  $Y$  is calculated from absorbance changes at any wavelength as follows,

$$Y = \frac{A - A_{\text{deoxy}}}{A_{\text{oxy}} - A_{\text{deoxy}}} \quad (4.1)$$

Here,  $A_{\text{oxy}}$  and  $A_{\text{deoxy}}$  are the absorbance of the fully oxygenated sample and the fully deoxygenated sample, respectively, and  $A$  is the absorbance at a given  $p\text{O}_2$ . This method is based on an assumption that the degree of change in absorbance of a haemoglobin solution at any fixed wavelength is proportional to  $Y$ , i.e. the ratio of oxygenated haems to total haems in the solution. Although this assumption has been accepted empirically, and there is no doubt about its validity for practical use, its strict validity is subject to question for certain reasons (refer to § 4.5.1). Dual-wavelength spectrophotometry is used for turbid samples such as whole blood or red cell suspensions.

Fig. 4.1 Light absorption spectra of oxyhaemoglobin (—), deoxyhaemoglobin (---), and aquomethaemoglobin (···) in the visible and near ultraviolet regions.  $E_{\text{mM}}$  is milimolar extinction coefficient in 0.1 M potassium phosphate buffer (pH 7.4).

